

Identifyn Standards for Optical Microscopy Methods™Identifyn Primary Antibody**Mouse anti-Human PARP1 Monoclonal, IgG1 - (5F10/B7)***Cat ID# - PA-MM-000001*Identifyn® Secondary Antibodies**Identifyn® SRM Alexa Fluor™ 647 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L)***Cat ID# - SA 000063*

PARP1 (Poly(ADP-ribose) polymerase 1) has a multifaceted function in the detection of DNA damage, including single-strand breaks (SSBs) and double-strand breaks (DSBs) via binding at the site of damage. PARP1 catalyzes the synthesis of poly(ADP-ribose) (PAR) chains using NAD⁺; the resulting PAR chains signal the recruitment of DNA repair proteins to the break, initiating the repair process. PARP1, as well as PARP2, are retained in nucleoli via interaction with the multifunctional nucleolar hub protein nucleophosmin, which is implicated in multiple steps of ribosome biogenesis, including rDNA transcription and elongation, as well as rRNA processing.

Method

Hela cells were grown on a µ-Slide 8 Well high Glass Bottom [Ibidi chamber](#), followed by -20°C methanol fixation and blocking with [Thermo Fisher Fish Serum Blocker \(cat#37527X3\)](#). The primary antibody, Mouse anti-Human PARP1 Monoclonal, was incubated for 1 hour at room temperature at a dilution of 1:100. Cells were washed 3X in Identifyn's™ proprietary Wash Buffer, and Identifyn® 647 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:50. Cells were washed 5X in Identifyn's™ proprietary Wash Buffer followed by the addition of Identifyn's™ proprietary STORM Buffer to the sample chamber enabling image acquisition.

Microscope

Bruker Vutara VXL Biplane Single Molecule Microscope, using an Olympus 60X/1.30 silicon oil objective, was used with the following parameters:

Calibration -

Calibration follows Identifyn's Calibration Standard Operating Procedure, which can be found [here](#).

<https://identifyn.com/storm-calibration>

Image Acquisition -

SRX Software - Under the 'Single Molecule Localization' tab, '**Record**' is selected and contains the following workflow tabs and parameters.

Configuration Tab - Capture Configuration

Number of Probes: 1

Number of Simultaneous Probes: 1
Number of Reference Probes: 1
Camera: Both Cameras
Color Channel: Both Channels
Probe Capture Mode: Interleaved
Z -Stack Capture Mode: Interleaved
Multiple Location Capture: Not selected
Post Record Reference: Not Selected
Fluidics: Not Selected
Autofocus Drift Correction: On

Microscope Configuration

Biplane Module: 635nm Dichroic
Objective: Silicon Selected
Stage Insert: Rectangular Selected
Course Focus Position: 8-Well Chamber Selected
Dichroic: SuperRes
PSF: (Red, Orange)
Heater: Current: 26

Camera Configuration

Recording Vertical Field of View: 100%
Widefield Camera Exposure Time: 100ms
Super Resolution Camera Exposure Time: 20ms
Widefield Camera Binning: 1
Per-Probe Exposure Time: Selected

Probe Configuration

Predefined Probe: First probe: 640 nm (635 nm Dichroic); First reference probe: 640 nm (635 nm Dichroic)
Photoactivation/Imaging Laser: 0% Selected
Exposure Time: 20ms Selected

Experiment Configuration

Record Workflow: Selected Options
View Mode: Widefield 200µm
Capture Mode: Live

Stage and Piezo Selection

Examination of the Sample

Cell(s) or Cellular Structure Localization - The red-bordered box identifies the area of the SuperRes view and defines a suitable cell(s) or cellular structure. The cell(s) or structures are optimally focused.

Piezo Current: 180.9 µm

Probe 1 Laser control 640 nm (635 nm Dichroic)

Widefield 640nm: 0.1X Neutral Density Filter: Selected, @0.2% Laser Input
Reference Probe 1 Laser control 640 nm (640 nm Dichroic): Not Selected

Advanced Tools Option

Widefield camera Exposure Time: 100ms
Default values are maintained unless otherwise noted.

Widefield Laser: "Widefield 640nm" was deselected, and "calibrate" under Autofocus was selected.

Autofocus - Calibration Values

Current Position" -0.979
Current Intensity: 4.5
Calibration Value: -5070 nm/unit
Autofocus Mode: Pre-Capture

View Mode

SuperRes: 50µm is selected

Capture Mode

Live

Stage and Piezo Selection

Cell(s) or Cellular Structure Localization - Positioning of the target Cell or Subsection of the Target Cell is optimized and optimally focused.

Piezo Record: Begin: 176.6; End: 178.6

Probe 1 Laser control 640 nm (635 nm Dichroic)

SuperRes 640nm: 0.1X Neutral Density Filter: Selected, @0.02% Laser Input

Reference Probe 1 Laser control 640 nm (635 nm Dichroic)

Not Selected

Advanced Tools

Widefield Camera Exposure Time: 20ms
Default values are maintained unless otherwise noted.

Record Workflow - Record Tab: Capture Parameters

Frames: Probe 1: 200
Z Positions: 4 using Steps: 0.2µm (200nm)

Cycles: 20
Time Points: 1
Pause Between Cycles: Not Selected
Automatic Photoactivation Laser Control: Not Selected

Probe 1 Laser control 640 nm (640 nm Dichroic)

Laser Power: 15%

Expert Configuration Options - The Localization Tab

Beyond the values shown, all other parameters were set to their default values.

Localization

Cutout Width: 16px
Max Frame Accumulation: 1 frame
Max Accumulation Offset: 1px
Fiducial at Max Accumulation: Not selected
Max Particles Per Frame: No Max
Background SLPE Removal: Not Selected
Cutout/Background Removal: Not Selected
Preview Cutout Positions: Selected

Record Workflow - Localization Tab**Data Collection**

The folder is created automatically with the localization file for this image acquisition.

Recording Summary

All parameters were carried over from previous tabs and are kept except:

Probe1: BG threshold: 12

Region of Interest -

Not selected

Display

Was deployed to inspect if the BG threshold applied chooses localizations in an expected manner

Record Workflow - View Tab**Region of Interest**

Default values

Visualization Parameters

Point Splatting
Sizing Mode: Radial Precision
Particle size: 4nm
Color by: Constant
Color Table: Single
Opacity Mode: Constant; Opacity 1: 03
Front-to-back Attenuation: Not Selected

Post Acquisition analysis - View Tab

Drift Correction - Computed with the following parameters:

Method: Particle (direct)
Time Intervals: 7 Intervals
Pixel Size: 50nm
Smoothing: 8.00

Advanced Particle Filters - Applied to the drift-corrected data with the following parameters

PSF Z offset: -600 to +600
Radial precision: 30
Axial precision: 70

The other parameters were not applied.

Statistical analysis - NA

Image by S. Thakur

© Copyright 2026 GMD12 LLC.