Identifyn Standards for Optical Microscopy Methods™

Identifyn Primary Antibody

Rabbit anti-Human 53BP1 Polyclonal, IgG
Cat ID# - PA-RP 000008

Identifyn® Secondary Antibodies

Identifyn®SRM Alexa Fluor™ 647 Affinity Purified Goat Anti-Rabbit, IgG (H + L)
Cat ID# - SA 000057

53BP1 (P53-binding protein 1) is a crucial player in the DNA damage response (DDR), particularly in the repair of double-strand breaks (DSBs). It has a range of functions that contribute to genomic stability and appropriate response to DNA damage. The following are the prominent roles and functions of 53BP1 in DNA repair:

DNA Damage Sensing: 53BP1 is among the early proteins recruited to sites of DNA damage. It recognizes and binds to the phosphorylated form of H2AX (γ-H2AX), a histone variant phosphorylated at double-strand break sites. This ability to localize to DNA damage sites helps to recruit other proteins involved in DNA repair and signaling.

Non-Homologous End Joining (NHEJ): 53BP1 significantly promotes NHEJ, a pathway to repair double-strand breaks, especially in the G1 phase of the cell cycle when a homologous template is unavailable. 53BP1 promotes this pathway by:

Facilitating Synapsis: 53BP1 helps align the broken DNA ends to prepare them for ligation by the NHEJ machinery.

Restricting End Resection: By inhibiting the end resection process (a step in homologous recombination), 53BP1 ensures that NHEJ remains the primary repair mechanism for DSBs in cells not in the S or G2 phase.

Interaction with Other DNA Repair Proteins: 53BP1 interacts with other proteins to promote NHEJ, such as RIF1 (Rap1 Interacting Factor 1) and the Shieldin complex. These interactions help determine the repair pathway choice, favoring NHEJ over homologous recombination when appropriate.

DNA Damage Signaling: 53BP1 is involved in amplifying DNA damage signaling, contributing to the recruitment of additional repair proteins and activating cell cycle checkpoints. This signaling function helps maintain genomic stability by ensuring that cells do not progress through the cell cycle with unrepaired DNA damage.

Tumor Suppression and Genome Stability: 53BP1 has tumor suppressive properties, given its role in promoting accurate repair and maintaining genomic stability. Its ability to ensure proper repair of DSBs and its interaction with other DDR proteins make it an essential player in preventing mutations and chromosomal aberrations that can lead to cancer.

In summary, 53BP1 is a multifaceted protein in the DNA damage response, promoting non-homologous end joining, facilitating DNA damage signaling, and contributing to genomic stability. Its interactions with other DNA repair proteins and its role in regulating the choice of repair pathways are crucial for maintaining cellular integrity and preventing tumorigenesis.

Method

© Copyright 2026 GMD12 LLC.

HeLa cells were grown on µ-Slide 8 Well high Glass Bottom Ibidi chamber. DNA damage was induced via treatment with 200µM Etoposide–spiked complete growth media for 12 hours, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Rabbit anti-Human 53BP1 Polyclonal was incubated for 1 hour at room temperature at a dilution of 1:100. Cells were washed 3X in Identifyn®'s proprietary Wash Buffer, and Identifyn®'s 647 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:100. Cells were washed 5X in Identifyn®'s proprietary Wash Buffer followed by the addition of Identifyn®'s proprietary STORM Buffer to the sample chamber enabling image acquisition.

Microscope

Bruker Vutara VXL Biplane Single Molecule Microscope, using an Olympus 60X/1.30 silicon oil objective, was used with the following parameters:

Calibration -

Calibration follows Identifyn®'s Calibration Standard Operating Procedure, which can be found here.

https://identifyn.com/storm-calibration

Image Acquisition -

SRX Software - Under the 'Single Molecule Localization' tab, **'Record'** is selected and contains the following workflow tabs and parameters.

Configuration Tab - Capture Configuration

Number of Probes: 1
Number of Simultaneous Probes: 1
Number of Reference Probes: 1
Camera: Both Cameras
Color Channel: Both Channels
Probe Capture Mode: Interleaved
Z -Stack Capture Mode: Interleaved
Multiple Location Capture: Not selected
Post Record Reference: Not Selected
Fluidics: Not Selected
Autofocus Drift Correction: On

Microscope Configuration

Biplane Module: 635nm Dichroic
Objective: Silicon Selected
Stage Insert: Rectangular Selected
Course Focus Position: 8-Well Chamber Selected
Dichroic: SuperRes
PSF: (Red, Orange)
Heater: Current: 26

© Copyright 2026 GMD12 LLC.

Camera Configuration

Recording Vertical Field of View: 100%
Widefield Camera Exposure Time: 100ms
Super Resolution Camera Exposure Time: 20ms
Widefield Camera Binning: 1
Per-Probe Exposure Time: Selected

Probe Configuration

Predefined Probe: First probe: 640 nm (635 nm Dichroic); First reference probe: 640 nm (635 nm Dichroic)
Photoactivation/Imaging Laser: 0% Selected
Exposure Time: 20ms Selected

Experiment Configuration

Record Workflow: Selected Options
View Mode: Widefield 200µm
Capture Mode: Live

Stage and Piezo Selection

Examination of the Sample

Cell(s) or Cellular Structure Localization - The red-bordered box identifies the area of the SuperRes view and defines a suitable cell(s) or cellular structure. The cell(s) or structures are optimally focused.

Piezo Current: 182.5 µm

Probe 1 Laser control 640 nm (635 nm Dichroic)

Widefield 640nm: 0.1X Neutral Density Filter: Selected, @0.2% Laser Input
Reference Probe 1 Laser control 640 nm (635 nm Dichroic): Not Selected

Advanced Tools Option

Widefield camera Exposure Time: 100ms
Default values are maintained unless otherwise noted.

Widefield Laser: "Widefield 640nm" was deselected, and "calibrate" under Autofocus was selected.

Autofocus - Calibration Values

Current Position: -0.979
Current Intensity: 4.5
Calibration Value: -5070 nm/unit
Autofocus Mode: Pre-Capture

View Mode

© Copyright 2026 GMD12 LLC.

SuperRes: 50µm is selected

Capture Mode

Live

Stage and Piezo Selection

Cell(s) or Cellular Structure Localization - Positioning of the target Cell or Subsection of the Target Cell is optimized and optimally focused.

Piezo Record: Begin: 182.5; End: 182.5

Probe 1 Laser control 640 nm (635 nm Dichroic)

SuperRes 640nm: 0.1X Neutral Density Filter: Selected, @0.02% Laser Input

Reference Probe 1 Laser control 640 nm (635 nm Dichroic)

Not Selected

Advanced Tools

Widefield Camera Exposure Time: 20ms
Default values are maintained unless otherwise noted.

Record Workflow - Record Tab: Capture Parameters

Frames: Probe 1: 30000
Z Positions: 1
Cycles: 1
Time Points: 1
Pause Between Cycles: Not Selected
Automatic Photoactivation Laser Control: Not Selected

Probe 1 Laser control 640 nm (635nm Dichroic)

Laser Power: 90%

Expert Configuration Options - The Localization Tab

Beyond the values shown, all other parameters were set to default values.

Localization

Cutout Width: 16px
Max Frame Accumulation: 1 frame
Max Accumulation Offset: 1px
Fiducial at Max Accumulation: Not selected
Max Particles Per Frame: No Max
Background SLPE Removal: Not Selected

© Copyright 2026 GMD12 LLC.

Cutout/Background Removal: Not Selected
Preview Cutout Positions: Selected

Record Workflow - Localization Tab

Data Collection

The folder is created automatically with the localization file for this image acquisition.

Recording Summary

All parameters were carried over from previous tabs and are kept except:

Probe1: BG threshold: 15

Region of Interest -

Not selected

Display

Was deployed to inspect if the BG threshold applied chooses localizations in an expected manner.

Record Workflow - View Tab

Region of Interest

Default values

Visualization Parameters

Point Splatting
Sizing Mode: Radial Precision
Particle size: 5nm
Color by: Depth
Color Table: Single
Opacity Mode: Depth; Opacity 1: 0.12
Front-to-back Attenuation: Not Selected

Post Acquisition analysis - View Tab

Drift Correction - Computed with the following parameters:

Method: Particle (direct)
Time Intervals: 10 Intervals
Pixel Size: 50nm
Smoothing: 8.00

Advanced Particle Filters - Applied to the drift-corrected data with the following parameters

© Copyright 2026 GMD12 LLC.

PSF Z offset: -999.71 to +999.89
Radial precision: 30
Axial precision: 70

The other parameters were not applied.

Statistical analysis - Computed to clean low background signals.

Statistics Dialog - Cluster Analysis

Clustering Algorithm: DBSCAN

Parameters

Maximum Particle Distance: 0.1um
Maximum Particle Count to Form Cluster: 10
Computer Hulls: Not Selected
Compute Density Isosurfaces: Not Selected
Edit Settings: All Probes
Compute: Selected
Following Cluster Analysis Computation, under the Results Tab, ID Unassigned: Deselected

---

Image by P. Raut

© Copyright 2026 GMD12 LLC.